

6 LANGUAGES VOICE(EN/FR/PL/ES/TR/CN)
32 MEMORY
SCAN FUNCTION(REAL TIME [TEMPERATURE](#)
[BODY](#)/OBJECTS MODE

1. Range for human body: 35.5C-42.9C
2. Range for object: 0C-118C
3. Back-lit display
4. C/F reading available
5. Auto recall of last reading
6. Automatic switch off
7. [Battery level indication](#)
8. Deviation setting: 0C-3.0C
9. Buzzer for fever alarm

Applicable standard: EN 12470-5 and ASTM 1965-98

Measurement distance: 5CM-8CM

Accuracy for human body:

$\pm 0.3^{\circ}\text{C}/0.5^{\circ}\text{F}$ 35.5-42.9 $^{\circ}\text{C}$ (95.9-107.6 $^{\circ}\text{F}$)

Accuracy for object:

$\pm 0.3^{\circ}\text{C}/0.5^{\circ}\text{F}$ 32.0-35.4 $^{\circ}\text{C}$ (89.6-95.7 $^{\circ}\text{F}$)

Resolution: 0.1 $^{\circ}\text{C}/0.1^{\circ}\text{F}$

Operation condition: 10.0 $^{\circ}\text{C}$ -40.0 $^{\circ}\text{C}$ (50.0 $^{\circ}\text{F}$ -104.0 $^{\circ}\text{F}$)

Humidity \leq 80% non-condensing

Storage condition: -25.0 $^{\circ}\text{C}$ -55.0 $^{\circ}\text{C}$ (-13.0 $^{\circ}\text{F}$ -131.0 $^{\circ}\text{F}$)

Humidity \leq 90% non-condensing

Battery: 2*AA battery





15 Accurate Measure Temperature monitor baby temperature while sleep

Error less than or equal to $\pm 0.2^{\circ}\text{C}$

Is accurate measure

Time (min)	Hot (°C)	Full (°C)
0	36.5	36.5
5	37.5	36.0
10	36.0	37.0
15	37.0	36.5
20	36.5	37.5
25	37.5	36.0
30	36.0	37.0
35	37.0	36.5
40	36.5	37.5
45	37.5	36.0

Website: thermometersupplier.com

Email: sanieleung@dor-medex.com